



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON NJ 08628-0068
(609) 882-2000

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

JEFFREY S. CHIESA
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

The New Jersey State Police (NJSP), Criminal Justice Records Bureau can no longer produce the Alcotest Memory Data Download Disc pertaining to this case. The Alcotest Centralized Database has been finalized and is available to the public. The database may be accessed from a link on the NJSP website, at [1. www.njportal.com/NJSP/Alcotest/](http://www.njportal.com/NJSP/Alcotest/). If you have questions about the Alcotest Program or data-related questions, contact the NJSP Office of Forensic Science; Phone: 609-584-5054 ex 5726 or via E-mail: OFSAlcotest@gw.njsp.org



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



Packing Slip

Customer no Order No Order date
 150063255 10855904 10/18/2016

Packing slip no. Ship date
 81032985 10/18/2016
 Please reference on inquiries

Ship to
 150063256
 TOWNSHIP OF HOLMDEL POLICE DEPARTME
 ATTN: CHRISTOPHER CHERNEY
 MONMOUTH COUNTY
 4 CRAWFORDS CORNER ROAD
 HOLMDEL, NJ 07733-1908
 USA

Payer

Your Purchase Order No. is
 CHRISTOPHER CHERNEY
 Any questions? Please contact:
 ROSEMARIE HUMBER

Customer
 TOWNSHIP OF HOLMDEL
 ATTN: FINANCE DEPARTMENT
 4 CRAWFORDS CORNER ROAD
 MONMOUTH COUNTY
 HOLMDEL, NJ 07733
 USA

Your sales person
 CHRISTOPHER FAY WILD
 Phone: 865-344-1330
 Fax: 972-917-1151

Delivery terms
 FA
 FREIGHT ALLOWED

Item#	Qty	Description
		Ship via: Fed Ex Saver
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 865-5900.
		REF: ANXB-0066 RECD 9/20/16
0010	1 EA	4414161 INFORMATION, 7110 ordered / Back ordered : 0 / 0 EA
0020	1 EA	4414155 MOTOR : PUEL USLL ordered / Back ordered : 0 / 0 EA

Draeger Safety Diagnostics Inc.
 Accounting Address
 101 Technology Blvd
 Pittsburgh, PA 15253
 Tel: (412) 789-5111
 Fax: (412) 789-5111

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6205426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 536410
 Pittsburgh, PA 15253-5906

Packing Slip

Customer no. Order date
150032255 10/18/2016

Packing slip no. Ship date
81032985 10/18/2016

Please reference on inquiries

Page 2 / 1

Item#	Sh. Quant	Description			
0430	0.5 EA	442-166 DAILY LABOR CHARGE Back ordered : 0.0 / 0.0 EA			

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Holmdel P.D., Holmdel, NJ

Date Received: 09/21/2016 Date given to service: 09/21/2016

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR XB-0066
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: November 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W

Repair Notes: Replaced defective fuel cell.

Fuel Cell = ARJE-2043
 CAL W/QC AND OPS CHECK

Service Technician BC Date: 10/18/2016

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Holmdel Police Department 4 Crawfords Corner Road Holmdel, NJ 07733		2. Contact: Christopher Cherney	4. Date: 09/15/2016
		3. Phone Number: 732-946-4400	5. County: Monmouth
6. Alcotest Instrument Serial Number: ARXB-0066	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
Complaint received of instrument experiencing control test failures.

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed low Electrochemical Sensor (EC) results (Fuel Cell) on control tests.

Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.


Please replace Fuel Cell.

Last Known Sequential File #: 01089

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II David Klimik #7040
 Name & Badge Number (Print)


 Signature

9/15/2016
 Date

Packing Slip

Customer no Order No Order date
 150063255 10685725 03/07/2014

Packing slip no. Ship date
 80841703 03/07/2014

Please reference on inquiries

ship to

150063256
 TOWNSHIP OF HOLMDEL POLICE DEPARTME
 ATTN: SGT. MICHAEL PIGOTT
 MONMOUTH COUNTY
 4 CRAWFORDS CORNER ROAD
 HOLMDEL, NJ 07733
 USA

bill to

Your Purchase Order Number

7110: ARXB-0066
 Any questions? Please contact:
 MARA HERRERA

Customer
 TOWNSHIP OF HOLMDEL
 ATTN: FINANCE DEPARTMENT
 4 CRAWFORDS CORNER ROAD
 MONMOUTH COUNTY
 HOLMDEL, NJ 07733
 USA

Your sales person

Don Pouliot
 Phone: 973-398-3228

Delivery terms
 FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARXB-0066			
0010	1 EA	MPCAL71 CALIBRATION CHARGE - 7110/9510 Ordered / Back ordered : 0 / 0 EA			
0020	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150063255 10685725 03/07/2014

Packing slip no. Ship date
80841703 03/07/2014

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6808486 PLATES,7110--9510 Ordered / Back ordered : 0 / 0 EA			
0040	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Holmdel Twp. P.D., Holmdel, NJ

Date Received: 02/28/14/ Date given to service: 02/28/14

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR XB-0066
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = AREL-1338
 CAL W/QC AND OPS CHECK

Service Technician BC Date: 03/07/2014

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Holmdel Twp. Police Dept. 4 Crawfords Corner Road Holmdel, NJ 07733		2. Contact: Sgt. Robert Philhower	4. Date: 2/6/2014
		3. Phone Number: (732) 946-4400	5. County: Monmouth
6. Alcotest Instrument Serial Number: ARXB-0066	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service.
At above location to perform a calibration and noticed the EC numbers on the Part I Control Test were low and believed the Fuel Cell is depleting.

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed a few repeated control test failures due to low Electrochemical Sensor (EC) results. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Sgt. Marc Dennis #5925
 Name & Badge Number (Print)

SGT.  **2/6/14**
 Signature Date

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Holmdel Twp. Police Dept. 4 Crawfords Corner Road Holmdel, NJ 07733	2. Contact: Sgt. Michael Pigott	4. Date: 9/20/11
	3. Phone Number: (732) 946-4400 Ext. 1731	5. County: Monmouth
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARXB-0066</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>

9. Reason for Service:
Returned from Draeger

10. Comments:
On 9/8/11 I was contacted by Joe Eckart from Draeger and he advised that he was able to clear the error message and that there was still memory intact. We then did a remote download, so I could save the digital downloads to disc. On 9/20/11 the Instrument was put back into service and a Calibration was performed on same with positive results.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Marc Dennis #5925

Name & Badge Number (Print)


 Signature

9/20/11
 Date

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Holmdel Twp P.D., Holmdel, NJ

Date Received: 09/07/11 Date given to service: _____

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR XB-0066
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>mp cal 71</u>	<u>Calibration</u>	<u>1</u>	<u>n/c w</u>
<u>mp labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/c w</u>

Repair Notes: ERROR 112 removed, Data down loaded by Supervisor.

Cal w/DC & ops check

Service Technician: [Signature] Date: 9-8-11

Packing Slip

Customer no Order No Order date
 150063255 10499504 09/12/2011

Packing slip no. Ship date
 80627178 09/12/2011
 Please reference on inquiries

ship to
 150063256
 TOWNSHIP OF HOLMDEL POLICE DEPARTME
 ATTN: SGT. KENNETH MCGOWAN
 MONMOUTH COUNTY
 4 CRAWFORDS CORNER ROAD
 HOLMDEL, NJ 07733
 USA

bill to
 [Empty Box]

Your Purchase Order Number
 7110 WARRANTY: ARXB-0066
 Any questions? Please contact:
 MARA HERRERA

Customer
 TOWNSHIP OF HOLMDEL
 ATTN: FINANCE DEPARTMENT
 P.O. BOX 410
 MONMOUTH COUNTY
 HOLMDEL, NJ 07733
 USA

Your sales person
 Don Pouliot
 Phone: 973-398-3228

Delivery terms
 FA
 FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARXB-0066			
0010	1 EA	MPCAL71 CALIBRATION CHARGE - 7110/9510 Ordered / Back ordered : 0 / 0 EA			
0020	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT


1. Department: Holmdel Twp. Police Dept. 4 Crawfords Corner Road Holmdel, NJ 07733	2. Contact: Sgt. Michael Pigott	4. Date: 9/1/11
	3. Phone Number: (732) 946-4400 Ext. 1731	5. County: Monmouth
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARXB-0066</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>

9. Reason for Service:
Complaint received of instrument displaying Error message 112 (Calibration)

10. Comments:
Operational condition of the instrument was assessed, however the error message remained displayed on the LED even after cycling the instrument on and off. No functions would work and the message could not be cleared from the instrument. The Instrument was placed out of service and returned to Draeger Safety Diagnostics for evaluation.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Marc Dennis #5925 Name & Badge Number (Print)	 Signature	9/1/11 Date
--	---	-----------------------